

A thick dark green vertical bar runs down the left side of the page. A dark green arrow points to the right, overlapping the bar, with the text "2020 Product List" inside it.

2020 Product List

Diagnostic Automation / Cortez Diagnostic Inc.

ELISA

Several thin, dark green wavy lines originate from the bottom left corner and curve upwards and to the right.

DA/CD Inc.

ELISA							
Product Name	Catalog #	# of Tests	CE	IVD	Qualitative / Quantitative	Assay Time	Sensitivity/ Specificity
Allergy							
AccuDiag™ Total Human IgE	1801-16	96	✓	IVD	Quantitative	95 min	5 IU/mL /98.7%
	1801-18 Min 5 kits	96	✓	IVD	Quantitative	80 min	5 IU/mL / 95%
AccuDiag™ Human Specific IgG 88 Food Allergens (New)	5154-8	96	✓	IVD	Quantitative	150 min	N/A
AccuDiag™ Human Specific IgG4 88 Food Allergens (New)	5155-8	96	✓	IVD	Quantitative	150 min	N/A
Allergy Line Blot Immunoassay							
AccuDiag™ Human Specific IgE (20 Seasonal Allergens) (New)	5150-8	16	✓	IVD	Qualitative	135 min	N/A
AccuDiag™ Human Specific IgE (20 Atopic Allergens) (New)	5151-8	16	✓	IVD	Qualitative	135 min	N/A
AccuDiag™ Human Specific IgE (20 Food Allergens) (New)	5152-8	16	✓	IVD	Qualitative	135 min	N/A
AccuDiag™ Human Specific IgE (20 Perennial Allergens) (New)	5153-8	16	✓	IVD	Qualitative	135 min	N/A
Autoimmune Diseases and Other							
AccuDiag™ ANA Screen	5102-2 Min 2 kits	96	✓	IVD	Qualitative	165 min	96.6%/100%
AccuDiag™ ENA Profile-6 (Sm/RNP, Sm, Jo-1, Scl-70, SS-A, SS-B)	2506-1 Min 2 kits	96	✓	IVD	Qualitative	80 min	> 96.7% / > 98.4%
	2506-2 Min 2 kits	96	✓	IVD	Semi Quantitative	65 min	> 92.0% / > 97.9%
AccuDiag™ ENA Combined Screen (6 antigens) Sm/RNP, Sm, Jo-1, Scl-70, SS-A, SS-B	2552-1 Min 2 kits	96	✓	IVD	Qualitative	80 min	96%/100%
	2552-2 Min 2 kits	96	✓	IVD	Qualitative	65 min	93/100%
AccuDiag™ Sm/RNP	2558-2 Min 2 kits	96	✓	IVD	Semi Quantitative	60 min	92/97.9%
AccuDiag™ RNP (New)	2510-1 Min 10 kits	96	✓	IVD	Semi Quantitative	90 min	95.6/96.9
AccuDiag™ Sm	2557-2 Min 2 kits	96	✓	IVD	Semi Quantitative	60 min	100/99.3%

ELISA							
Product Name	Catalog #	# of Tests	CE	IVD	Qualitative / Quantitative	Assay Time	Sensitivity/ Specificity
AccuDiag™ Jo-1	2570-2 Min 2 kits	96	✓	IVD	Semi Quantitative	60 min	100/100%
AccuDiag™ Scl-70	2559-2 Min 2 kits	96	✓	IVD	Semi Quantitative	60 min	100/100%
AccuDiag™ SS-A (Ro)	2555-2 Min 2 kits	96	✓	IVD	Semi Quantitative	60 min	98.2/100%
AccuDiag™ SS-B (La)	2556-2 Min 2 kits	96	✓	IVD	Semi Quantitative	60 min	96.6/100%
AccuDiag™ dsDNA	2553-1 Min 2 kits	96	✓	IVD	Semi Quantitative	65 min	100/97.9%
AccuDiag™ Anti-Phospholipid IgG/IgM	2560-6 Min 2 kits	96	✓	IVD	Quantitative	65 min	IgG: 91.8%/97.3% IgM: 53.4%/96.7%
	2560-17 Min 5 kits	96	✓	IVD	Quantitative	125 min	IgG: 92.3%/84.6% IgM: 68.8%/99.9%
AccuDiag™	1491-2 Min 2 kits	96	✓	IVD	Semi Quantitative	60 min	91.7%/75.9%
AccuDiag™ Cardiolipin IgM	1492-2 Min 2 kits	96	✓	IVD	Semi Quantitative	60 min	93.6%/76.5%
AccuDiag™ Cardiolipin IgA	1493-2 Min 2 kits	96	✓	IVD	Semi Quantitative	60 min	94.4%/77%
AccuDiag™ Cardiolipin IgG, A, M	1490-1 Min 10 kits	96	✓	IVD	Qualitative	120 min	5.2 EU/ml
AccuDiag™ Beta 2 Glycoprotein 1 IgA	1494-11	96	✓	IVD	Quantitative	75 min	98/91%
AccuDiag™ Beta 2 Glycoprotein 1 IgG	1495-11	96	✓	IVD	Quantitative	75 min	91/92%
AccuDiag™ Beta 2 Glycoprotein 1 IgM	1496-11	96	✓	IVD	Quantitative	75 min	92.6/85.4%
AccuDiag™ Histone	2571-1 Min 2 kits	96	✓	IVD	Semi Quantitative	75 min	98.8/90.9%
AccuDiag™ Rheumatoid Factor (RF) IgM	2550-1 Min 2 kits	96	✓	IVD	Qualitative	65 min	100/98.7%
AccuDiag™ Mitochondria	1420-1 Min 2 kits	96	✓	IVD	Semi Quantitative	75 min	100/100%
AccuDiag™ Thyroglobulin	1421-1 Min 2 kits	96	✓	IVD	Semi Quantitative	75 min	100/96.7%
	1421-2 Min 2 kits	96	✓	IVD	Qualitative & Semi Quantitative	60 min	100%/100%
AccuDiag™ TBG (Thyroid Binding Globulin) New!	3177-15	96	✓	IVD	Quantitative		1ug/ml/100%

ELISA							
Product Name	Catalog #	# of Tests	CE	IVD	Qualitative / Quantitative	Assay Time	Sensitivity/ Specificity
AccuDiag™ TPO (Thyroid Peroxidase)	1446-2 Min 2 kits	96	✓	IVD	Semi Quantitative	60 min	93.7/94.4%
AccuDiag™ PR3 (c-ANCA)	1445-1	96	✓	IVD	Semi Quantitative	75 min	98.2/95.6%
	1445-2 Min 2 kits	96	✓	IVD	Qualitative & Semi Quantitative	60 min	90.3%/98.6%
AccuDiag™ ANCA Screen IgG	1681-2 Min 2 kits	96	✓	IVD	Semi Quantitative	90 min	97.4/94.9%
AccuDiag™ MPO, Myeloperoxidase (p-ANCA)	1441-2 Min 2 kits	96	✓	IVD	Qualitative & Semi Quantitative	60 min	98.9/97.9%
AccuDiag™ Gliadin IgG	1021-2 Min 2 kits	96	✓	IVD	Qualitative & Semi Quantitative	60 min	96.6/95.6%
AccuDiag™ Gliadin IgA	1022-2 Min 2 kits	96	✓	IVD	Qualitative & Semi Quantitative	60 min	93.9/96.2%
AccuDiag™ Anti CCP	3175-17 Min 5 kits	96	✓	IVD	Quantitative	110 min	97/79%
AccuDiag™ Anti-Tissue Transglutaminase IgA	3173-17 Min 5 kits	96	✓	IVD	Quantitative	80 min	88.5/100%
AccuDiag™ Anti-Tissue Transglutaminase IgG	3174-17 Min 5 kits	96	✓	IVD	Quantitative	110 min	81/90.9%
Anemia							
AccuDiag™ Folate	3127-15	96	✓	IVD	Quantitative	95 min	0.52 ng/ml / N/A
AccuDiag™ Ferritin	1601-16	96	✓	IVD	Quantitative	110 min	5 ng/ml / 98.5%
	1601-18 Min 5 kits	96	✓	IVD	Quantitative	65 min	5 ng/ml / 96.5%
AccuDiag™ Vitamin B12	3125-15	96	✓	IVD	Quantitative	80 min	70.13 pg/ml
AccuDiag™ sTfR - Soluble Transferrin Receptor	3126-15	96	✓	IVD	Quantitative	105 min	0.055 nmol/L /96%
Bone Metabolism							
AccuDiag™ Intact PTH	1735-6	96	✓	IVD	Quantitative	250 min	1.57 pg/mL
	1735-15	96	✓	IVD	Quantitative	105	0.49 pg/ml

ELISA

Product Name	Catalog #	# of Tests	CE	IVD	Qualitative / Quantitative	Assay Time	Sensitivity/ Specificity
AccuDiag™ 25-OH Vitamin D (total)	2065-6	96	✓	IVD	Quantitative	85 min	>2.89 ng/mL / 100%
	2065-15	96	✓	IVD	Quantitative	80 min	1.2 ng/ml
Cancer (Tumor Marker)							
AccuDiag™ Prostatic Acid Phosphatase (PAP)	4227-11	96	✓	IVD	Quantitative	75 min	0.2 ng/ml / 98.7%
AccuDiag™ Beta 2-Microglobulin	4200-16	96	✓	IVD	Quantitative	95 min	2.0 µg/mL / 96%
	4200-18 Min 5 kits	96	✓	IVD	Quantitative	80 min	0.1 µg/mL / 93%
AccuDiag™ PSA	4222-15	96		In-Vitro	Quantitative	75 min	0.013 ng/ml / 99%
	4222-16	96		In-Vitro	Quantitative	155 min	0.5 ng/ml / 98.7%
AccuDiag™ AFP (Alpha Fetoprotein)	5101-16	96	✓	IVD	Quantitative	95 min	2.0 ng/ml / 98.5%
AccuDiag™ CEA	5201-16	96	✓	IVD	Quantitative	95 min	1.0 ng/ml / 95%
AccuDiag™ CA-15-3	6333-16	96	✓	IVD	Quantitative	235 min	5U/mL/97.5 %
AccuDiag™ CA-12-5	6502-16	96	✓	IVD	Quantitative	215 min	5U/ml / 97%
AccuDiag™ CA-19-9	6909-16	96	✓	IVD	Quantitative	155 min	5U/ml / 97%
AccuDiag™ NSE	6334-16	96	✓	IVD	Quantitative	110 min	15 ng/ml / 98.7%
AccuDiag™ Free Beta HCG	4221-16	96	✓	IVD	Quantitative	110 min	0.5 mIU/ml / 100%
	4221-18 Min 5 kits	96	✓	IVD	Quantitative	80 min	0.25 ng/ml / -
AccuDiag™ Free PSA	4322-16	96		In-Vitro	Quantitative	180 min	0.05ng/ml / 98.7%
AccuDiag™ Pro-GRP (Gastrin-Releasing Peptide) (New)	7015-49	96	✓	IVD	Semi-Quantitative	120 min	N/A
AccuDiag™ S100B (New)	1792-17 Min 5 kits	96	✓	IVD	Quantitative	210 min	35.27 pg/mL

ELISA

Product Name	Catalog #	# of Tests	CE	IVD	Qualitative / Quantitative	Assay Time	Sensitivity/ Specificity
Cardiac Markers							
AccuDiag™ Digoxin	1669-15	96	✓	IVD	Quantitative	75 min	0.072 ng/ml
AccuDiag™ CK-MB	1005-15	96	✓	IVD	Quantitative	60 min	0.05 ng/ml / 96%
AccuDiag™ Troponin I	1105-18 Min 5 kits	96	✓	IVD	Quantitative	125 min	1 ng/ml / 96%
AccuDiag™ Myoglobin	1667-18 Min 5 kits	96	✓	IVD	Quantitative	80 min	5 ng/ml / 96%
AccuDiag™ LP (a)	1740-1 Min 2 kits	96	✓	IVD	Quantitative	100 min	0.8 mg/dL/100%
AccuDiag™ C-Reactive Protein (CRP)	1668-18 Min 5 kits	96	✓	IVD	Quantitative	80 min	0.01 mg/ml / 96%
Diabetes and Pancreatic Disease Assay							
AccuDiag™ Insulin	1606-15	96	✓	IVD	Quantitative	165 min	0.182 µIU/ml / 100%
AccuDiag™ C-Peptide	1293-15	96	✓	IVD	Quantitative	165 min	0.02 ng/ml / 100%
AccuDiag™ Leptin	1742-6	96	✓	IVD	Quantitative	205 min	1 ng/ml / N/A
AccuDiag™ Adiponectin (New)	1743-38	96	✓	IVD	Quantitative	105 min	0.06 ng/ml / 100%
AccuDiag™ Anti-GAD (New)	1791-17 Min 5 kits	96	✓	IVD	Quantitative	160 min	62.3%/88.6%
AccuDiag™ IAA (New)	1790-17 Min 5 kits	96	✓	IVD	Quantitative	105 min	77%/94%
Fertility							
AccuDiag™ Beta hCG (Total)	4201-16	96	✓	IVD	Quantitative	110 min	2 mIU/ml / 97%
AccuDiag™ Human Growth Hormone (HGH)	1901-16	96	✓	IVD	Quantitative	110 min	1 ng/mL
	1901-18 Min 5 kits	96	✓	IVD	Quantitative	65 min	0.5 ng/mL
AccuDiag™ hCG Visual	4102-16	96	✓	IVD	Quantitative	30 min	20 MIU/ml / 96%
AccuDiag™ FSH	4224-16	96	✓	IVD	Quantitative	95 min	2 mIU/ml / 95%

ELISA

Product Name	Catalog #	# of Tests	CE	IVD	Qualitative / Quantitative	Assay Time	Sensitivity/ Specificity
	4224-18 Min 5 kits	96	✓	IVD	Quantitative	65 min	2.5 mIU/ml / -
AccuDiag™ LH	4225-16	96	✓	IVD	Quantitative	95 min	2 mIU/ml / 96%
	4225-18 Min 5 kits	96	✓	IVD	Quantitative	65 min	1 mIU/ml / 94.5%
AccuDiag™ PAPP-A	4229-6	96	✓	IVD	Quantitative	85 min	0.133 µg/ml / 97%
	4229-15	96	✓	IVD	Quantitative	110 min	3.12 mIU/ml
AccuDiag™ Prolactin (PRL)	4226-16	96	✓	IVD	Quantitative	80 min	2 ng/ml / 96%
AccuDiag™ SHBG	2043-6	96	✓	IVD	Quantitative	75 min	0.77 nmol/L / 100%
	2043-17 Min 5 kits	96	✓	IVD	Quantitative	75 min	0.2 nmol/mL

FOOD (The minimum Order is 5 Kits each or in Total)

AccuDiag™ Deoxynivalenol (DON) (New)	5147-8	96	N/A	FOOD	Quantitative	90 min	0.5 ng/mL / N/A
AccuDiag™ Zearalenone (New)	5148-8	96	N/A	FOOD	Quantitative	230 min	0.25 ng/mL / N/A
AccuDiag™ T2-Toxin (New)	5149-8	96	N/A	FOOD	Quantitative	50 min	0.9-117 ppb/ N/A
AccuDiag™ Vitamin H (Biotin)	1741-8	96	N/A	FOOD	Quantitative	90 min	0.5 ng/mL / N/A
AccuDiag™ Vitamin B12	1739-8	96	N/A	FOOD	Quantitative	80 min	0.3 ng/mL / N/A
AccuDiag™ Folate (Folic Acid)	1736-8	96	N/A	FOOD	Quantitative	140 min	2 ng/mL / N/A
AccuDiag™ Campylobacter Antigen Detection	8321-3	96	N/A	FOOD	Qualitative	50 min	N/A
AccuDiag™ E. Coli 0157 Antigen Detection	8322-3	96	N/A	FOOD	Qualitative	70 min	N/A
AccuDiag™ Verotoxin Antigen Detection	8328-3	96	N/A	FOOD	Qualitative	100 min	See insert
AccuDiag™ Egg White Protein	5138-8	96	N/A	FOOD	Quantitative	60 min	0.05 ppm
AccuDiag™ Beta-Lactoglobulin (Milk Protein)	5136-8	96	N/A	FOOD	Quantitative	60 min	1.5 ppb

ELISA

Product Name	Catalog #	# of Tests	CE	IVD	Qualitative / Quantitative	Assay Time	Sensitivity/ Specificity
AccuDiag™ Soy Protein	5144-8	96	N/A	FOOD	Quantitative	60 min	16 ppb
AccuDiag™ Aflatoxin B1	5120-8	96	N/A	FOOD	Quantitative	140 min	5 ng/mL
AccuDiag™ Aflatoxin M1	5121-8	96	N/A	FOOD	Quantitative	140 min	10 pg/mL
AccuDiag™ Aflatoxin Total (New)	5156-8	96	N/A	FOOD	Quantitative	45 min	See Insert
AccuDiag™ Fumonisin	5123-8	96	N/A	FOOD	Quantitative	80 min	0.5 ng/ml
AccuDiag™ Chloramphenicol	5126-8	96	N/A	FOOD	Quantitative	60 min	0.03 ng/mL
AccuDiag™ Penicillin	5127-8	96	N/A	FOOD	Quantitative	140 min	3 ng/mL
AccuDiag™ Streptomycin	5128-8	96	N/A	FOOD	Quantitative	60 min	1 ng/ml
AccuDiag™ Histamine	5133-8	96	N/A	FOOD	Quantitative	80 min	2 ng/ml
AccuDiag™ Almond	5135-8	96	N/A	FOOD	Quantitative	60 min	0.25 ng/mL
AccuDiag™ Gluten / Gliadin	5139-8	96	N/A	FOOD	Quantitative	60 min	0.3 ppm
AccuDiag™ Hazelnut	5140-8	96	N/A	FOOD	Quantitative	60 min	0.33 ppm
AccuDiag™ Lupine	5141-8	96	N/A	FOOD	Quantitative	60 min	0.2 ppm
AccuDiag™ Mustard	5142-8	96	N/A	FOOD	Quantitative	60 min	1 ppm
AccuDiag™ Peanut	5143-8	96	N/A	FOOD	Quantitative	60 min	0.1 ppm
AccuDiag™ Ochratoxin A	1743-8	96	N/A	FOOD	Quantitative	15 min	0.1 – 1.6 ppb
AccuDiag™ Walnut	5145-8	96	N/A	FOOD	Quantitative	60 min	0.35 ppm
AccuDiag™ Casein (New)	5157-8	96	N/A	FOOD	Quantitative	60 min	0.04 ppm
AccuDiag™ Cashew (New)	5158-8	96	N/A	FOOD	Quantitative	60 min	0.2 ppm
AccuDiag™ Coconut (New)	5160-8	96	N/A	FOOD	Quantitative	60 min	0.4 ppm

ELISA

Product Name	Catalog #	# of Tests	CE	IVD	Qualitative / Quantitative	Assay Time	Sensitivity/ Specificity
AccuDiag™ Crustaceans (New)	5161-8	96	N/A	FOOD	Quantitative	60 min	See insert
AccuDiag™ Fish (New)	5162-8	96	N/A	FOOD	Quantitative	60 min	1.4 ppm
AccuDiag™ Lysozyme (New)	5163-8	96	N/A	FOOD	Quantitative	60 min	2 ppb
AccuDiag™ Macademia (New)	5164-8	96	N/A	FOOD	Quantitative	60 min	0.1 ppm
AccuDiag™ Milk (New)	5165-8	96	N/A	FOOD	Quantitative	60 min	0.05 ppm
AccuDiag™ Molluscs (New)	5166-8	96	N/A	FOOD	Quantitative	60 min	See insert
AccuDiag™ Pecan Nut (New)	5167-8	96	N/A	FOOD	Quantitative	60 min	0.2 ppm
AccuDiag™ Pistachio (New)	5168-8	96	N/A	FOOD	Quantitative	60 min	0.13 ppm
AccuDiag™ Sesame (New)	5169-8	96	N/A	FOOD	Quantitative	60 min	0.2 ppm

Infectious Diseases

AccuDiag™ Chikungunya IgG	8112-25	96	✓	IVD	Qualitative	100 min	100/ 95%
AccuDiag™ Chikungunya IgM	8113-25	96	✓	IVD	Qualitative	100 min	100 / 100%
AccuDiag™ Dengue IgG/IgM	8115-35	96	✓	IVD	Qualitative	25 min	100/100%
AccuDiag™ Dengue IgG	8116-35	96	✓	IVD	Qualitative	25 min	100/96%
AccuDiag™ Dengue IgM	8117-35	96	✓	IVD	Semi Quantitative	40 min	93.5 / 97.8 %
AccuDiag™ Dengue NS1 Antigen	8404-25	96	✓	IVD	Qualitative	110 min	95 / 95%
AccuDiag™ Adenovirus (Fecal)	8300-3	96	✓	IVD	Qualitative	45 min	100 / 93.1 %
AccuDiag™ Anti-Tetanus Toxoid	8900-28	96	✓	IVD	Qualitative	90 min	0.0093 IU/ml /97%
AccuDiag™ Diphtheria Toxoid IgG	1760-28	96	✓	IVD	Quantitative	90 min	0.006 IU/mL /97%
AccuDiag™ HPV IgG	1889-6	96	✓	IVD	Semi-quantitative	140 min	N/A

ELISA							
Product Name	Catalog #	# of Tests	CE	IVD	Qualitative / Quantitative	Assay Time	Sensitivity/ Specificity
AccuDiag™ West Nile IgG	8400-25	96	✓	IVD	Qualitative	15 min	96.2/98.4 %
AccuDiag™ West Nile IgM	8401-25	96	✓	IVD	Qualitative	190 min	96.2/98.4 %
AccuDiag™ Rotavirus (Fecal)	8306-3	96	✓	IVD	Qualitative	45 min	100/ 97.1%
AccuDiag™ Japanese Encephalitis IgG	8402-25	96	✓	IVD	Qualitative	135 min	N/A
AccuDiag™ Japanese Encephalitis IgM	8403-25	96	✓	IVD	Qualitative	196 min	100/ 100%
AccuDiag™ Bordetella Pertussis IgA	5117-8 Min 3 kits	96	✓	IVD	Qualitative	170 min	94 / 92%
AccuDiag™ Bordetella Pertussis IgG	5118-8 Min 3 kits	96	✓	IVD	Qualitative	170 min	86 / 88%
AccuDiag™ Bordetella Pertussis IgM	5119-8 Min 3 kits	96	✓	IVD	Quantitative	170 min	100 / 88%
AccuDiag™ RSV IgA	5113-8 Min 5 kits	96	✓	IVD	Qualitative	170 min	100/ 100%
AccuDiag™ RSV IgG	5114-8 Min 3 kits	96	✓	IVD	Qualitative	170 min	86/ 100%
AccuDiag™ RSV IgM	5115-8 Min 3 kits	96	✓	IVD	Qualitative	170 min	86/ 100%
AccuDiag™ Adenovirus IgA	6103-8 Min 3 kits	96	✓	IVD	Qualitative	170 min	100/ 97.7 %
AccuDiag™ Adenovirus IgG	6111-8 Min 3 kits	96	✓	IVD	Qualitative	170 min	100/ 100%
AccuDiag™ Adenovirus IgM	6102-8 Min 3 kits	96	✓	IVD	Qualitative	170 min	100/ 100%
AccuDiag™ Influenza A IgA	5104-8 Min 3 kits	96	✓	IVD	Qualitative	170 min	83 / 99%
AccuDiag™ Influenza A IgG	5105-8 Min 3 kits	96	✓	IVD	Qualitative	170 min	100 / 88%
AccuDiag™ Influenza A IgM	5106-8 Min 3 kits	96	✓	IVD	Qualitative	170 min	100/ 100%
AccuDiag™ Influenza B IgA	5107-8 Min 3 kits	96	✓	IVD	Qualitative	170 min	85/ 100%
AccuDiag™ Influenza B IgG	5108-8 Min 3 kits	96	✓	IVD	Qualitative	170 min	100 / 89%
AccuDiag™ Influenza B IgM	5109-8 Min 3 kits	96	✓	IVD	Qualitative	170 min	100/ 100%

ELISA

Product Name	Catalog #	# of Tests	CE	IVD	Qualitative / Quantitative	Assay Time	Sensitivity/ Specificity
AccuDiag™ Mycobacterium tuberculosis (TB) IgA	5110-8 Min 3 kits	96	✓	IVD	Qualitative	170 min	100 / 99%
AccuDiag™ Mycobacterium tuberculosis (TB) IgG	5111-8 Min 3 kits	96	✓	IVD	Qualitative	170 min	100 / 99%
AccuDiag™ Mycobacterium tuberculosis (TB) IgM	5112-8 Min 3 kits	96	✓	IVD	Qualitative	170 min	100/ 100%
AccuDiag™ Toxoplasma IgG	1101-1 Min 2 kits	96		In-Vitro	Quantitative	60 min	95.3 / 100%
	1101-2 Min 2 kits	96		In-Vitro	Qualitative/ Quantitative	60 min	93.8% / 96.1%
AccuDiag™ Toxoplasma IgM	1102-1 Min 2 kits	96		In-Vitro	Qualitative	75 min	100/ 100%
	1102-2 Min 2 kits	96		In-Vitro	Qualitative	95 min	92.9% / 100%
AccuDiag™ Toxoplasma IgA	1103-11	96		In-Vitro	Qualitative	75 min	94 / 100%
AccuDiag™ Rubella IgG	1301-1 Min 2 kits	96x2		IVD	Qualitative	65 min	
	1301-2 Min 2 kits	96		In-Vitro	Qualitative/ Quantitative	95 min	100% / 100%
AccuDiag™ Rubella IgM	1302-2 Min 2 kits	96		In-Vitro	Qualitative	60 min	100% / 97.2%
	1302-18	96		In-Vitro	Qualitative	90 min	100% / 99.2%
AccuDiag™ Rubella IgA	1300-11	96		In-Vitro	Qualitative	75 min	100/ 99.2%
AccuDiag™ Cytomegalovirus IgG (CMV IgG)	1201-1 Min 2 kits	96x2		IVD	Semi-Quantitative	65 min	
	1201-2 Min 2 kits	96		In-Vitro	Qualitative	70 min	96.4 / 93.9%
AccuDiag™ Cytomegalovirus IgM (CMV IgM)	1202-2 Min 2 kits	96		In-Vitro	Qualitative	70 min	92.5 / 98.5%
AccuDiag™ Cytomegalovirus IgA (CMV IgA)	1200-11	96		In-Vitro	Qualitative	75 min	99.2 / 97.2%
AccuDiag™ Herpes Simplex 1 IgG (HSV1 IgG)	1401-11	96	✓	IVD	Qualitative	75 min	100/ 100%
AccuDiag™ Herpes Simplex 1 IgM (HSV1 IgM)	1402-11	96	✓	IVD	Qualitative	75 min	75 / 96%
AccuDiag™ Herpes Simplex 1 IgA (HSV1 IgA)	1400-11	96	✓	IVD	Qualitative	90 min	98 / 98.7%

ELISA

Product Name	Catalog #	# of Tests	CE	IVD	Qualitative / Quantitative	Assay Time	Sensitivity/ Specificity
AccuDiag™ Herpes Simplex 2 IgG (HSV2 IgG)	1403-11	96	✓	IVD	Qualitative	75 min	100/98%
AccuDiag™ Herpes Simplex 2 IgM (HSV2 IgM)	1404-11	96	✓	IVD	Qualitative	75 min	99/100%
AccuDiag™ AccuDiag™ Herpes Simplex 2 IgA (HSV2 IgA)	1399-11	96	✓	IVD	Qualitative	75 min	98 / 98.7%
AccuDiag™ Herpes Simplex 1,2 IgG (HSV1,2 IgG)	1443-11	96	✓	IVD	Qualitative	75 min	100/ 100%
AccuDiag™ Herpes Simplex 1,2 IgM (HSV1,2 IgM)	1444-2 Min 2 kits	96	✓	IVD	Qualitative	65 min	75% / 90.3%
AccuDiag™ Epstein-Barr Viral Capsid Antigen (P-18) IgG	1405-1 Min 2 kits	96x2	✓	IVD	Qualitative	65 min	95% (sero+) /100%
AccuDiag™ Epstein-Barr Virus VCA IgG (EBV, VCA IgG)	1405-2 Min 2 kits	96	✓	IVD	Qualitative	60 min	98.1% / 98.3%
AccuDiag™ Epstein-Barr Virus VCA IgA EBV, VCA IgA) (EBV, VCA IgA)	1414-11	96	✓	IVD	Qualitative	75 min	100/ 100%
AccuDiag™ Epstein-Barr Virus VCA IgM (EBV, VCA IgM)	1406-1 Min 2 kits	96x2	✓	IVD	Qualitative	65 min	
	1406-2 Min 2 kits	96	✓	IVD	Qualitative	90 min	90.9% / 93.2%
AccuDiag™ Epstein-Barr Virus Early Antigen (EA) IgG	1415-1 Min 2 kits	96x2	✓	IVD	Qualitative	65 min	
	1415-2 Min 2 kits	96	✓	IVD	Qualitative	75 min	92% / 81%
AccuDiag™ Epstein-Barr Virus Early Antigen (EA) IgM	1418-16	96	✓	IVD	Qualitative	75 min	100/ 100%
AccuDiag™ Epstein-Barr Virus Nuclear Antigen (EBNA) IgG	1425-1 Min 2 kits	96	✓	IVD	Semi-Quantitative	75 min	97.8/ 100%
AccuDiag™ Epstein-Barr Virus Nuclear Antigen (EBNA) IgM	1426-1 Min 2 kits	96	✓	IVD	Qualitative	135 min	100/94.5 %
AccuDiag™ Chlamydia Trachomatis IgG	1407-11	96		In-Vitro	Qualitative	75 min	96.1/98.5%
AccuDiag™ Chlamydia Trachomatis IgA	1417-11	96		In-Vitro	Qualitative	75 min	100/94%
AccuDiag™ Chlamydia Trachomatis IgM	1427-11	96		In-Vitro	Qualitative	75 min	99/99%
AccuDiag™ Measles IgG	1408-1 Min 2 kits	96	✓	IVD	Quantitative	75 min	99.3 / 91 %
AccuDiag™ Measles IgM	1409-1 Min 2 kits	96	✓	IVD	Qualitative	75 min	93/93%

ELISA							
Product Name	Catalog #	# of Tests	CE	IVD	Qualitative / Quantitative	Assay Time	Sensitivity/ Specificity
AccuDiag™ Mumps IgG	1410-1 Min 2 kits	96	✓	IVD	Quantitative	75 min	99.3 / 96.6%
AccuDiag™ Mumps IgM	1411-1 Min 2 kits	96	✓	IVD	Qualitative	75 min	100 / 100%
AccuDiag™ Legionella Total (IgG/M/A)	1651-2 Min 2 kits	96	✓	IVD	Qualitative	105 min	92.3 / 93.1%
AccuDiag™ Legionella IgG/IgM	1652-1 Min 2 kits	96	✓	IVD	Qualitative	75 min	90 / 98%
AccuDiag™ H. pylori IgG	1503-2 Min 2 kits	96	✓	IVD	Qualitative	60 min	95% / 100%
AccuDiag™ H. pylori IgA	1505-11	96	✓	IVD	Qualitative	75 min	88.9 / 97.10%
AccuDiag™ H. pylori IgM	1504-11	96	✓	IVD	Qualitative	75 min	98 / 98%
AccuDiag™ H. pylori Antigen	1506-11	96	✓	IVD	Quantitative. 5 ng/mL /	75 min	0.5 ng/ml / N/A
AccuDiag™ Varicella-Zoster IgG	1412-1 Min 2 kits	96	✓	IVD	Quantitative. 5 ng/mL /	75 min	99.4 / 97%
AccuDiag™ Varicella-Zoster IgM	1413-1 Min 2 kits	96	✓	IVD	Qualitative	50 min	100 / 100%
AccuDiag™ Lyme Disease IgG (B. burgdorferi IgG)	1428-2 Min 2 kits	96	✓	IVD	Qualitative	100 min	91.7 / 96.4 %
AccuDiag™ Lyme Disease IgM (B. burgdorferi IgM)	1424-2 Min 2 kits	96	✓	IVD	Qualitative	105 min	95.5 / 92.5 %
AccuDiag™ Lyme Disease IgG/IgM (B. burgdorferi IgG,IgM)	1423-1 Min 2 kits	96x2	✓	IVD	Qualitative	65 min	See insert
	1423-2 Min 2 kits	96	✓	IVD	Qualitative	105 min	92 / 95 %
AccuDiag™ Mycoplasma pneumoniae IgG	8042-2 Min 2 kits	96	✓	IVD	Qualitative	105 min	95 / 87.5 %
	8042-1 Min 2 kits	96x2	✓	IVD	Semi-Quantitative	65 min	See insert
AccuDiag™ Mycoplasma pneumoniae IgM	8043-1 Min 2 kits	96x2	✓	IVD	Semi-Quantitative	65 min	See insert
	8043-2 Min 2 kits	96	✓	IVD	Qualitative	105 min	100 / 93 %
AccuDiag™ Brucella IgG	1501-8 Min 3 kits	96	✓	IVD	Quantitative	170 min	100 / 100%
AccuDiag™ Brucella IgM	1502-8 Min 3 kits	96	✓	IVD	Qualitative	170 min	100 / 100%

ELISA

Product Name	Catalog #	# of Tests	CE	IVD	Qualitative / Quantitative	Assay Time	Sensitivity/ Specificity
AccuDiag™ Brucella IgA	1507-8 Min 3 kits	96	✓	IVD	Qualitative	170 min	100 / 100%
AccuDiag™ Parvovirus B19 IgG	1450-2 Min 2 kits	96	✓	IVD	Qualitative	60 min	5-7 IU/mL/100%
AccuDiag™ Parvovirus B19 IgM	1451-2 Min 2 kits	96	✓	IVD	Qualitative	60 min	See insert
The minimum Order for Hepatitis Panel, HIV, and Syphilis should be 10Kits of each product or in total THE NEW DEVELOPED HEPATITIS PANEL ELISA IS COMING SOON							
AccuDiag™ HBsAg	1701-12	96		In-Vitro	Qualitative	85 min	99.9 / 99.75%
AccuDiag™ HBsAg Ultra-Sensitive	1703-12	96		In-Vitro	Qualitative	130 min	100/ 99.78 %
AccuDiag™ HBsAb	1702-12	96		In-Vitro	Qualitative	75 min	100 / 99.70 %
AccuDiag™ HBsAb Quantitative	1710-12	96		In-Vitro	Quantitative	85 min	5 mIU/ml / 99%
AccuDiag™ HBcAb	1704-12	96		In-Vitro	Qualitative	105 min	99.92 / 99.82%
AccuDiag™ HBcAb IgM	1778-12	96		In-Vitro	Qualitative	75 min	98.4 / 99.3%
AccuDiag™ HBeAg	1705-12	96		In-Vitro	Qualitative	105 min	100 / 95%
AccuDiag™ HBeAb	1706-12	96		In-Vitro	Qualitative	105 min	100 / 97.7 %
AccuDiag™ HDV IgG	1870-12	96	✓	IVD	Qualitative	85 min	100 / 100%
AccuDiag™ HDV IgM Discontinued	1869-12	96	✓	IVD	Qualitative	75 min	100 / 100%
AccuDiag™ HEV IgG	1888-12	96	✓	IVD	Qualitative	100 min	93.8 / 85.4 %
AccuDiag™ HEV IgM	1876-12	96	✓	IVD	Qualitative	110 min	97.1 / 100 %
AccuDiag™ HEV Ab	1877-12	96	✓	IVD	Qualitative	100 min	99.8 / 99.6 %
AccuDiag™ HCV Ab	1707-12	96		In-Vitro	Qualitative	105 min	100 / 99.55 %
AccuDiag™ HAV IgG	1519-12	96	✓	IVD	Qualitative	105 min	100 / 100%
AccuDiag™ HAV IgM	1718-12	96	✓	IVD	Qualitative	75 min	100 / 99 %

ELISA

Product Name	Catalog #	# of Tests	CE	IVD	Qualitative / Quantitative	Assay Time	Sensitivity/ Specificity
AccuDiag™ HIV 1/2	1516-12	96		In-Vitro	Qualitative	90 min	99.95 / 99.89%
AccuDiag™ HIV 1/2 Ab/Ag	1520-12	96		In-Vitro	Qualitative	135 min	100 / 100 %
AccuDiag™ HTLV 1 + 2 Discontinued	8196-12	96	✓	IVD	Qualitative	120 min	100 / 99.99 %
AccuDiag™ Syphilis (TPA) IgG	1463-1 Min 2 kits	96	✓	IVD	Qualitative	180 min	100 / 100%
AccuDiag™ Syphilis (TPA) IgM	1464-6	96	✓	IVD	Qualitative	135 min	95 / 95%
AccuDiag™ Syphilis (Anti-TP) IgG/IgM	1465-12	96	✓	IVD	Qualitative	90 min	N/A / 100%
AccuDiag™ Salmonella typhi IgG	6150-47	96	✓	IVD	Qualitative	50 min	N/A
AccuDiag™ Salmonella typhi IgM	6151-47	96	✓	IVD	Qualitative	50 min	N/A

NeoNatal Panel

AccuDiag™ NTSH (NeoNatal TSH)	3124-15	192	✓	IVD	Quantitative	95 min	0.2 µIU/mL / 100%
AccuDiag™ NNT4 (NeoNatal T4)	3150-15	96	✓	IVD	Quantitative	155 min	0.05 µg/dL / 97%
AccuDiag™ N-TBG (NeoNatal TBG) (New)	4115-15	96	✓	IVD	Quantitative	135 min	0.60 ng/ml
AccuDiag™ N-17OHPG (Neo Natal 17-OH progesterone) (New)	4116-15	96	✓	IVD	Quantitative	135 min	28.6 nmol/L

Parasitology Antibody Detection

AccuDiag™ Cysticercosis IgG (T.Solium)	8105-35	96	✓	IVD	Semi-quantitative	20 min	88 / 96%
AccuDiag™ E. histolytica IgG (Amebiasis)	8201-35	96	✓	IVD	Qualitative	20 min	92 / 100 %
AccuDiag™ Echinococcus IgG	8202-35	96	✓	IVD	Qualitative	20 min	97.9 / 91.7 %
AccuDiag™ Leishmania	8203-25 Min 30 kits	96	✓	IVD	Qualitative	65 min	100% / 99%
	8203-35	96	✓	IVD	Qualitative	25 min	97 / 84 %
AccuDiag™ Fasciola IgG	8119-35	96	✓	IVD	Qualitative	20 min	100 / 100 %

ELISA							
Product Name	Catalog #	# of Tests	CE	IVD	Qualitative / Quantitative	Assay Time	Sensitivity/ Specificity
AccuDiag™ Fasciola gigantica (New)	8120-35	96	✓	IVD	Qualitative	20 min	See insert
AccuDiag™ Schistosoma IgG	8209-35	96	✓	IVD	Qualitative	25 min	100 / 85 %
AccuDiag™ Strongyloides IgG	8319-35	96	✓	IVD	Qualitative	20 min	100 / 100 %
AccuDiag™ Chagas	8100-35	96	✓	IVD	Qualitative	20 min	96 / 97 %
	8100-33	96	✓	IVD	Qualitative	90 min	100/98.7%
AccuDiag™ Scrub typhus IgG	8406-25	96	✓	IVD	Qualitative	75 min	N/A
AccuDiag™ Scrub typhus IgM	8405-25	96	✓	IVD	Qualitative	75 min	N/A
AccuDiag™ Leptospira IgG	8204-35	96		IVD	Qualitative	85 min	80%/87.5%
AccuDiag™ Leptospira IgM	8208-35	96	✓	IVD	Qualitative	85 min	100 / 100 %
AccuDiag™ Leptospira IgG & IgM	8210-35	96	✓	IVD	Qualitative	20 min	88.7 / 87.5 %
AccuDiag™ Toxocara IgG	8206-35	96	✓	IVD	Qualitative	80 min	87.5 / 93.3 %
AccuDiag™ Trichinella IgG	8207-35	96	✓	IVD	Qualitative	80 min	100 / 100 %
AccuDiag™ Ascaris IgG	8212-35	96	✓	IVD	Qualitative	20 min	100 / 96 %
AccuDiag™ Filaria IgG4 (New)	8211-25	96	✓	IVD	Qualitative	60 min	N/A
Parasitology Antigen Detection							
AccuDiag™ Cryptosporidium (Fecal)	8301-3	96	✓	IVD	Qualitative	105 min	100 / 100 %
AccuDiag™ E. Coli 0157 (Fecal)	8302-3	96	✓	IVD	Qualitative	130 min	100 / 92 %
AccuDiag™ E. histolytica/dispar	8307-3	96	✓	IVD	Qualitative	45 min	88 / 100 %
AccuDiag™ E. Coli Verotoxin (Fecal)	8303-3	96	✓	IVD	Qualitative	100 min	100 / 99 %
AccuDiag™ Giardia	8304-3	96	✓	IVD	Qualitative	105 min	100 / 100 %

ELISA

Product Name	Catalog #	# of Tests	CE	IVD	Qualitative / Quantitative	Assay Time	Sensitivity/ Specificity
AccuDiag™ Crypto/Giardia Ag Combo	8310-3	96	✓	IVD	Qualitative	105 min	99 / 100 %
AccuDiag™ Legionella Urinary Antigen	8305-3	96	✓	IVD	Qualitative	45 min	100 / 99.2 %
AccuDiag™ Campylobacter (Fecal)	8320-3	96	✓	IVD	Qualitative	50 min	77 / 100 %
Steroids							
AccuDiag™ Testosterone Saliva	7095-38	96	✓	IVD	Quantitative	125 min	1 pg/ml / 100%
AccuDiag™ Cortisol Saliva	7101-38	96	✓	IVD	Quantitative	80 min	1.0 ng/ml / 100%
AccuDiag™ Cortisol	6101-15	96	✓	IVD	Quantitative	105 min	0.37 µg/ml/ 95%
AccuDiag™ Estradiol	2046-18 Min 5 kits	96	✓	IVD	Quantitative	125 min	10 pg/ml / 97%
AccuDiag™ DHEA-S	2055-15	96	✓	IVD	Quantitative	110 min	0.03 µg/ml/ 90%
	2055-17 Min 3 kits	96	✓	IVD	Quantitative	75 min	0.03 µg/ml/ 90%
AccuDiag™ DHEA	2056-6	96	✓	IVD	Quantitative	75 min	0.07 ng/ml / Ins.
	2056-15	96	✓	IVD	Quantitative	80 min	0.10 ng/ml / Ins.
AccuDiag™ Aldosterone	4230-6	96	✓	IVD	Quantitative	75 min	10.0 pg/ml / 100%
AccuDiag™ Progesterone	2077-18 Min 5 kits	96	✓	IVD	Quantitative	125 min	0.06 ng/mL / 97%
AccuDiag™ Testosterone	2095-18 Min 5 kits	96	✓	IVD	Quantitative	125 min	0.05 ng/mL / 95%
AccuDiag™ Free Testosterone	2925-15	96	✓	IVD	Quantitative	105 min	N/A
	2925-17 Min 3 kits	96	✓	IVD	Quantitative	75 min	0.06 pg/ml / 100%
AccuDiag™ Androstenedione	1038-17 Min 3 kits	96	✓	IVD	Quantitative	105 min	0.01 ng/mL / 100%
AccuDiag™ Estrone (New)	2066-38	96	✓	IVD	Quantitative	95 min	3pg/ml / 100%
AccuDiag™ Free Estriol	3171-15	96	✓	IVD	Quantitative	110 min	0.03 ng/ml / 100%

ELISA

Product Name	Catalog #	# of Tests	CE	IVD	Qualitative / Quantitative	Assay Time	Sensitivity/ Specificity
	3171-17 Min 3 kits	96	✓	IVD	Quantitative	75 min	0.03 ng/ml / 100%
AccuDiag™ Total Estriol (New)	2067-17 Min 3 kits	96	✓	IVD	Quantitative	75 min	1.05 ng/ml / 100%
AccuDiag™ 17-alpha-OH Progesterone	1292-15	96	✓	IVD	Quantitative	110 min	0.05 ng/ml / 100%
AccuDiag™ Dihydro-Testosterone (DHT)	2042-38	96	✓	IVD	Quantitative	95 min	6.0 pg/ml / 100%

Thyroid (Endocrinology)

AccuDiag™ T3	3144-16	96	✓	IVD	Quantitative	95 min	0.2ng/mL / 96.3%
	3144-18 Min 5 kits	96	✓	IVD	Quantitative	80 min	0.2ng/mL / 96.8%
AccuDiag™ T4	3149-16	96	✓	IVD	Quantitative	80 min	0.5 µg/dl / 96.3%
	3149-18 Min 5 kits	96	✓	IVD	Quantitative	80 min	0.5 µg/dl / 96%
AccuDiag™ TSH	3122-16	96	✓	IVD	Quantitative	80 min	0.2 µIU/mL / 100%
	3122-18 Min 5 kits	96	✓	IVD	Quantitative	80 min	0.2 µIU/ml / 97.5%
AccuDiag™ TSH Ultra-Sensitive	3123-16	96	✓	IVD	Quantitative	170 min	0.05 µIU/mL / 96%
	3123-18 Min 5 kits	96	✓	IVD	Quantitative	140 min	0.05 µIU/mL / 96%
AccuDiag™ Free T4	3146-15	96	✓	IVD	Quantitative	105 min	0.314 ng/dL / 97%
AccuDiag™ Free T3	3148-15	96	✓	IVD	Quantitative	105 min	0.05pg/mL / 97%
AccuDiag™ T3 Uptake	3176-15	96	✓	IVD	Quantitative	105 min	N/A
AccuDiag™ Reverse T3 (New)	3145-38	96	✓	IVD	Quantitative	105 min	0.009 ng/ml/100%

Others

AccuDiag™ ACTH (New)	2044-47	96	N/A	Research	Quantitative	5 hours	1 pg/ml
AccuDiag™ C1 Esterase (New)	1925-45	96	N/A	Research	Quantitative	180	1.875 ng/ml

ELISA

Product Name	Catalog #	# of Tests	CE	IVD	Qualitative / Quantitative	Assay Time	Sensitivity/ Specificity
AccuDiag™ Hepcidin 25 (bioactive) HS (New)	1744-6	96	N/A	Research	Quantitative	110 min	See insert
AccuDiag™ Thyroglobulin Cancer (New)	1421-15	96	CE	IVD	Quantitative	See insert	See insert

